

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program NOV 20 2009

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N	130180
Name		AW Anderson	ID# 2467 Date 11/16/09
A	Agency	Seldovia Police Dept.	Phone # (907) 234-7640
Instrument Location		Seldovia Police Dept.	
Alco S/N		X301530	Target Value 0.078 High Pressure 600 PSI
B	Alco Test Value Average	0.074	0.075
		1 st Alco	2 nd Alco
Signature		AW Anderson	
		(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 21st day of Dec, 2009.

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

BT9 01/00



of Datamaster cdm
State of Alaska
Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N 130180 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180 ✓

NOVEMBER 16, 2009

OPERATOR'S NAME: *CBS 12/1/09*
ANDERSON AM
OPERATOR'S NUMBER: 2467
SUBJECT'S LAST NAME:
VERIFICATION OF CALIBRATION
SUBJECT'S FIRST NAME/MI:
N/A
S.L. #: 1234567890
DEPT/AGENCY: 9041
CASE/REPORT: N/A
TEST TYPE: U
ALCO TARGET VALUE: .075 ✓
ALCO S/N: N201508 ✓

--- BREATH ANALYSIS ---

.075 ADJUSTED FOR 29.51 in ✓		
ALCO TARGET	.075	14:21
BLANK TEST	.000	14:22
INTERNAL STANDARD	VERIFIED	14:22
ALCO TO 29.51 in	.074 ✓	14:22
BLANK TEST	.000	14:23
SUBJECT SAMPLE	.000	14:25
BLANK TEST	.000	14:25
ALCO TO 29.51 in	.075 ✓	14:26
BLANK TEST	.000	14:27

CBS 12/1/09

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130180 ✓

NOVEMBER 16, 2009 ✓
TIME 14:18

--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
SOFTWARE DATE:	02/20/01
HEATERS	
SAMPLE CHAMBER:	49c
BREATH TUBE:	41c
BAROMETER:	29.51 in
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

11#12/01/09 14:27:00 (130180) 1234567890
HIJKLMNORSTUVWXYZ01234567890
qrstuvwxyz01234567890